

JAN 18 2011
2011
3/9/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130177</u> ✓
Name	<u>STEVEN C. KEVAN</u>	ID# <u>6585</u> ✓ Date <u>1/11/11</u> ✓
A	Agency <u>ALASKA STATE TROOPERS</u>	Phone # <u>907-438-2089</u>
	Instrument Location <u>ST. MARY'S</u>	
B	Alco S/N <u>X 172959</u> ✓	Target Value <u>1.081</u> ✓ High Pressure <u>800</u>
	Alco Test Values	<u>1.085</u> ✓ 1st Alco
		<u>1.085</u> ✓ 2nd Alco
	Signature <u>[Signature]</u>	
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 11th day of MAR, 2011 DN

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130177 SKZ

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

JANUARY 11, 2011

OPERATOR'S NAME:

KEVIN / STEVE

OPERATOR'S NUMBER: 6885

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: 00

DEPT/AGENCY: SMD

CASE/REPORT: 11-3223

TEST TYPE: U

ALCO TARGET VALUE: .081

ALCO S/N: K172959

— BREATH ANALYSIS —

.081 ADJUSTED FOR 30.64 in

ALCO TARGET .082 12:18

BLANK TEST .000 12:19

INTERNAL STANDARD VERIFIED 12:19

ALCO TV 30.64 in .085 12:19

BLANK TEST .000 12:20

SUBJECT SAMPLE .000 12:20

BLANK TEST .000 12:21

ALCO TV 30.64 in .085 12:21

BLANK TEST .000 12:22

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177

JANUARY 11, 2011

TIME 12:15

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.64 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ\]^_`abcdefg|hijklmno
pqrstuvwxyz{|}~■